

Schedule

Hylab Pte. Ltd.
63 Ubi Crescent
Singapore 408599

Certificate No. : LA-2013-0549-C
Issue No. : 9
Date : 12 December 2022
Page : 1 of 18

FIELD OF TESTING : Calibration and Measurement

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
A. ELECTRICAL (Lab and Site)		
1. Calibration of Sourcing Instrument: DC Voltage	In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter	
1 µV – 0.1 mV		0.8 µV
0.1 mV - 25 mV		1.1 µV
25 mV - 100 mV		2.1 µV
100 mV - 0.5 V		9.0 µV
0.5 V - 1 V		17 µV
1 V - 5 V		0.1 mV
5 V - 10 V		0.2 mV
10 V - 50 V		1.5 mV
50 V - 100 V		2.9 mV
100 V - 250 V		8.0 mV
250 V - 500 V		15 mV
500 V - 750 V		22 mV
750 V - 1000 V		29 mV
2. Calibration of Sourcing Instrument: AC Voltage	In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter	
1 mV)	50 Hz - 1kHz	4.6 µV
1 mV - 5 mV)		5.0 µV
5 mV - 10 mV)		5.6 µV

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 2 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
1 mV)	1 kHz - 20 kHz	4.6 μ V
1 mV - 5 mV)		5.2 μ V
5 mV - 10 mV)		6.4 μ V
1 mV)	20 kHz - 50 kHz	4.9 μ V
1 mV - 5 mV)		8.3 μ V
5 mV - 10 mV)		14 μ V
1 mV)	50 kHz - 100 kHz	9 μ V
1 mV - 5 mV)		31 μ V
5 mV - 10 mV)		60 μ V
10 mV - 50 mV)	50 Hz - 1 kHz	7.7 μ V
50 mV - 75 mV)		10 μ V
75 mV - 100 mV)		12 μ V
10 mV - 50 mV)	1kHz - 20 kHz	12 μ V
50 mV - 75 mV)		16 μ V
75 mV - 100 mV)		19 μ V
10 mV - 50 mV)	20 kHz - 50 kHz	21 μ V
50 mV - 75 mV)		29 μ V
75 mV - 100 mV)		38 μ V
10 mV - 50 mV)	50 kHz - 100 kHz	49 μ V
50 mV - 75 mV)		72 μ V
75 mV - 100 mV)		95 μ V
0.1 V - 1 V	50 Hz - 1 kHz	0.2 mV
0.1 V - 1 V	1 kHz - 20 kHz	0.2 mV
0.1 V - 1 V	20 kHz - 50 kHz	0.4 mV
0.1 V - 1 V	50 kHz - 100 kHz	1.0 mV

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 3 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
1 V - 2.5 V) 50 Hz - 1 kHz	0.5 mV
2.5 V - 5 V)	0.7 mV
5 V - 10 V)	1.1 mV
1 V - 2.5 V) 1 kHz - 20 kHz	0.7 mV
2.5 V - 5 V)	1.1 mV
5 V - 10 V)	1.9 mV
1 V - 2.5 V) 20 kHz - 50 kHz	2.1 mV
2.5 V - 5 V)	3.8 mV
5 V - 10 V)	3.8 mV
1 V - 2.5 V) 50 kHz - 100 kHz	2.7 mV
2.5 V - 5 V)	4.9 mV
5 V - 10 V)	9.5 mV
10 V - 25 V) 50 Hz - 20 kHz	8.5mV
25 V - 50 V)	15 mV
50 V - 100 V)	26 mV
100 V - 230 V) 50 Hz - 1 kHz	0.2 V
230 V - 700 V)	0.4 V
3. Calibration of Sourcing Instrument: DC Current	In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter	
1 µA - 10 µA		0.8 nA
10 µA - 75 µA		3.2 nA
75 µA - 100 µA		4.0 nA
0.1 mA -10 mA		0.4 µA
10 mA - 50 mA		3.0 µA
50 mA - 100 mA		6.0 µA
0.1 A - 0.5 A		0.1 mA
0.5 A - 1 A		0.2 mA
1 A - 10 A	Using current shunt	0.04 A
10 A - 20 A		0.07 A
20 A - 30 A		0.11 A

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 4 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
4. Calibration of Sourcing Instrument: AC Current	In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter	
10 µA - 100 µA	50 Hz - 5 kHz	0.2 µA
0.1 mA - 1 mA) 50 Hz - 500 Hz	1.0 µA
1 mA - 10 mA)	9.4 µA
10 mA - 50 mA)	60 µA
50 mA - 100 mA)	94 µA
0.1 A - 0.5 A)	0.9 mA
0.5 A - 1 A)	1.5 mA
0.1 mA - 1 mA) 500 Hz - 5 kHz	1 µA
1 mA - 10 mA)	6.4 µA
10 mA - 50 mA)	53 µA
50 mA - 100 mA)	68 µA
0.1 A - 0.5 A)	1 mA
0.5 A - 1 A)	1.5 mA
1 A - 10 A	50 Hz & 1 kHz	0.06 A
10 A - 20 A	Using Current Shunt	0.11 A
5. Calibration of Sourcing Instrument: Resistance, 2 wire	In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter	
0 mΩ - 10 Ω		0.6 Ω
10 Ω - 100 Ω		0.6 Ω
100 Ω - 1 kΩ		0.6 Ω
1 kΩ - 10 kΩ		0.8 Ω
10 kΩ - 100 kΩ		2.4 Ω
100 kΩ - 1 MΩ		0.1 kΩ
1 MΩ - 10 MΩ		1.1 kΩ
10 MΩ - 100 MΩ		0.12 MΩ
100 MΩ - 1 GΩ		12 MΩ
1 GΩ	In-house procedure	6.0 MΩ
2 GΩ	HYL/WI/E-07, Rev: 04,	12 MΩ
5 GΩ		32 MΩ

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 5 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
<p>6. Calibration of Sourcing Instrument: Resistance, 4 wire</p> <p>0 Ω - 1 Ω 1 Ω - 10 Ω 10 Ω - 100 Ω 100 Ω - 1 kΩ 1 kΩ - 10 kΩ 10 kΩ - 100 kΩ</p>	<p>In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter</p>	<p>0.2 mΩ 0.4 mΩ 2.5 mΩ 19 mΩ 0.2 Ω 1.8 Ω</p>
<p>7. Calibration of Sourcing Instrument: Frequency</p> <p>10 Hz - 25 Hz 25 Hz - 50 Hz 50 Hz - 100 Hz 100 Hz - 500 Hz 500 Hz - 1 kHz 1 kHz - 5 kHz 5 kHz - 10 kHz 10 kHz - 50 kHz 50 kHz - 100 kHz 100 kHz - 1000 kHz</p>	<p>In-house procedure HYL/WI/E-04, Rev: 02, 8½ digit multimeter</p>	<p>16 mHz 30 mHz 13 mHz 59 mHz 0.2 Hz 0.6 Hz 1.2 Hz 5.8 Hz 12 Hz 0.2 kHz</p>
<p>8. Calibration of Measuring Instrument: DC Voltage</p> <p>0 mV - 50 mV 50 mV - 200 mV 0.2 V - 1 V 1 V - 2 V 2 V - 10 V 10 V - 20 V 20 V - 100 V 100 V - 200 V 200 V - 750 V 750 V - 1000 V</p>	<p>In-house procedure HYL/WI/E-03, Rev: 01, Transmille 3041, Using Direct Method with 4-½ DMM</p>	<p>8.1 μV 16 μV 78 μV 0.13mV 0.8 mV 1.8 mV 7.9 mV 13 mV 73 mV 80 mV</p>

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 6 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
9. Calibration of Measuring Instrument: AC Voltage	In-house procedure HYL/WI/E-03 Rev: 01, Transmille 3041 Using Direct Method with 4-½ DMM	
1 mV - 100 mV	50 Hz 50 Hz – 10 kHz 10 kHz – 20 kHz	0.1 mV 0.3mV 0.9mV
100 mV - 200 mV	50 Hz 50 Hz – 10 kHz 10 kHz – 20 kHz	0.3 mV 0.5 mV 1.7 mV
0.2 V – 2 V	50 Hz 50 Hz – 10 kHz 10 kHz – 20 kHz	1.9 mV 3.6 mV 16 mV
2 V – 10 V	50 Hz 50 Hz – 10 kHz 10 kHz – 20 kHz	9.4 mV 20 mV 0.11 V
10 V – 20 V	50 Hz 50 Hz – 10 kHz 10 kHz – 20 kHz	18 mV 37 mV 0.16 V
20 V – 100 V	50 Hz 50 Hz – 20 kHz	98 mV 0.3 V
100 V – 200 V	50 Hz 50 Hz – 20 kHz	0.2 V 0.5 V
200 V – 1000 V	50 Hz 50 Hz – 10 kHz	1.0 V 4.1 V

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 7 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
10. Calibration of Measuring Instrument: DC Current	In-house procedure HYL/WI/E-03 Rev: 01, Transmille 3041	62 nA
0.2 μ A – 100 μ A	Using Direct Method with 4- $\frac{1}{2}$ DMM	75 nA
100 μ A – 200 μ A		0.8 μ A
0.2 mA – 2 mA		1.3 μ A
2 mA – 10 mA		2.0 μ A
10 mA – 20 mA		15 μ A
20 mA – 100 mA		23 μ A
100 mA - 200 mA		0.2 mA
0.2 A - 1 A		0.4 mA
1 A - 2 A		3.7 mA
2 A - 5 A		5.9 mA
5 A - 10 A		12 mA
10 A - 20 A		
10 mA – 100 mA	In-house procedures	0.4 mA
100 mA - 1000 mA	HYL/WI/E-03 Rev: 01,	3.3 mA
1 A - 10 A	HYL/WI/E-09 Rev: 01,	0.3 A
10 A - 100 A	Transmille 3041, Current	0.9 A
100 A - 200 A	Clamp EA002	1.9 A
200 A - 500 A		3.5 A
500 A - 1000 A		6.3 A
1000 A - 1200 A		7.6 A
11. Calibration of Measuring Instrument: AC Current	In-house procedure HYL/WI/E-03 Rev: 01, Transmille 3041, Using Direct Method with 4- $\frac{1}{2}$ DMM	
20 μ A - 200 μ A) 50 Hz - 500 Hz	0.8 μ A
0.2 mA - 1 mA)	3.2 μ A
1 mA - 2 mA)	3.9 μ A
2 mA - 10 mA)	31 μ A
10 mA - 20 mA)	38 μ A
20 mA - 200 mA)	0.4 mA
0.2 A - 2 A)	4.0 mA
2 A - 10 A)	27 mA
10 A – 20 A)	31 mA

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 8 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD / FREQUENCY	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
20 μ A - 200 μ A) 500 Hz - 1 kHz	4.0 μ A
0.2 mA - 1 mA)	12 μ A
1 mA - 2 mA)	21 μ A
2mA - 20 mA)	0.3 mA
20 mA - 200 mA)	1.4 mA
0.2 A - 2 A)	17 mA
2 A - 10 A)	54 mA
10 A - 20 A)	92 mA
10 mA – 100 mA	@ 50 Hz, In house procedures	0.4 mA
100 mA - 1000 mA	HYL/WI/E-03 Rev: 01,	3.5 mA
1 A - 10 A	HYL/WI/E-09 Rev: 01,	0.3 A
10 A - 100 A	Transmille 3041,	0.9 A
100 A - 200 A	Current Clamp EA002	1.9 A
200 A - 500A		3.6 A
500 A - 1000 A		6.2 A
1000 A - 1200 A		7.6 A
12. Calibration of Measuring Instrument: Resistance, 2 wire	In-house procedure HYL/WI/E-03, Rev: 01, Transmille 3041 Using Direct Method with 4-½ DMM	
0.1 Ω		47 m Ω
0.2 Ω		47 m Ω
1 Ω		47 m Ω
10 Ω		49 m Ω
100 Ω		54 m Ω
1 k Ω		0.2 Ω
10 k Ω		1.3 Ω
100 k Ω		13 Ω
1 M Ω		0.3 k Ω
10 M Ω		5.4 k Ω
100 M Ω		0.7 M Ω

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 9 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
<p>13. Calibration of Measuring Instrument: 4-wired Resistance</p> <p>0.01Ω 0.1 Ω 1 Ω 10 Ω 100 Ω 1 kΩ 10 kΩ 100 kΩ</p>	<p>In-house procedure HYL/WI/E-03, Rev 01, Transmille 3041 Using Direct Method with 4-½ DMM</p>	<p>5.8 mΩ 5.8 mΩ 9 mΩ 10 mΩ 15 mΩ 0.2 Ω 1.2 Ω 12 Ω</p>
<p>14. Calibration of Measuring Instrument: Frequency</p> <p>10 Hz – 100 Hz 100 Hz - 500 Hz 0.5 kHz - 5 kHz 5 kHz - 50 kHz 50 kHz - 500 kHz 500 kHz - 1000 kHz</p>	<p>In-house procedure HYL/ WI/E-03 Rev: 01, Transmille 3041 Using Direct Method with 4-½ DMM</p>	<p>6.7 mHz 18 mHz 0.2 Hz 1.7 Hz 20 Hz 34 Hz</p>
<p>15. Calibration of Measuring Instrument: Capacitance</p> <p>1 nF 10 nF 20 nF 50 nF 100 nF 1 μF 10 μF</p>	<p>In-house procedure HYL/WI/E-14, Rev: 01, Transmille 3041 Using Direct Method with 4-½ DMM</p> <p>) 1 kHz))))))</p>	<p>0.01 nF 0.09 nF 0.2 nF 0.3 nF 0.5 nF 0.004 μF 0.10 μF</p>

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 10 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)	
16. Inductance Measuring Instrument	In-house procedure HYL/WI/E-14, Rev: 01, Transmille 3041		
1 mH)		8.7 μ H
10 mH)		84 μ H
20 mH)		84 μ H
30 mH) 1 kHz		0.3 mH
50 mH)		0.5 mH
100 mH)		0.9 mH
1 H)		8.4 mH
10 H)		87 mH
17. Stopwatch and Digital Timer	In-house procedure HYL / WI / E-11, Rev: 02, Digital Timer	<u>Lab</u>	<u>On-Site</u>
10 s – 90 s		0.4 s	0.5 s
90 s - 480 s		0.6 s	0.6 s
480 s - 1800 s		2.3 s	2.4 s
1800 s - 3600 s		2.9 s	3.0 s
3600 s - 7200 s		5.1 s	5.2 s
18. Calibration of Measuring Instrument: Earth Resistance / Insulation Resistance test	In-house procedure HYL / WI / E-10, Rev: 01, Resistance Box		
Earth Resistance			
0.01 Ω - 0.1 Ω			0.0007 Ω
0.1 Ω - 1 Ω			0.001 Ω
1 Ω - 100 Ω			0.06 Ω
100 Ω - 1000 Ω			0.6 Ω
1 k Ω - 10 k Ω			6 Ω
10 k Ω - 20 k Ω			12 Ω
Insulation Resistance			
0.5 M Ω - 10 M Ω			0.12 M Ω
10 M Ω - 100 M Ω			1.2 M Ω
100 M Ω - 500 M Ω			13 M Ω
500 M Ω - 1000 M Ω			18 M Ω
1 G Ω - 2 G Ω			0.04 G Ω
2 G Ω - 5 G Ω			0.17 G Ω

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 11 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)															
<p>B. MECHANICAL</p> <p>1. Tachometer (Non-Contact) (Lab only)</p> <p>240 rpm – 900 rpm 900 rpm – 3000 rpm 3 000 rpm – 12 000 rpm 12 000 rpm – 18 000 rpm 18 000 rpm – 19 992 rpm 19 992 rpm – 30 000 rpm 30 000 rpm – 39 996 rpm 39 996 rpm – 49 992 rpm 49 992 rpm – 60 000 rpm</p> <p>2. Weights (Lab only)</p> <p>1 mg – 10 mg Above 10 mg – 500 mg Above 500 mg – 20 g Above 20 g – 100 g Above 100 g – 200 g Above 200 g – 500 g Above 500 g – 8.2 kg Above 8.2 kg – 20 kg</p> <p>3. Weighing Balance and Scales</p> <p>1 mg – 220 g Upto 1000 g Upto 8.2 kg Upto 30 kg Upto 150 kg Upto 250 kg</p>	<p>In-house procedure HYL / WI / M-20 Rev: 02</p> <p>In-house procedure HYL / WI / M-06 Rev: 02</p> <p>In-house procedure HYL / WI / M-11 Rev: 04</p>	<p>0.1 rpm 0.2 rpm 1.0 rpm 1.3 rpm 1.4 rpm 1.9 rpm 2.4 rpm 3.0 rpm 3.6 rpm</p> <p>10 µg 20 µg 30 µg 40 µg 0.12 mg 9 mg 10 mg 1 g</p>	<table border="1"> <thead> <tr> <th><u>Lab</u></th> <th><u>On-site</u></th> </tr> </thead> <tbody> <tr> <td>0.2 mg</td> <td>0.3 mg</td> </tr> <tr> <td>2 mg</td> <td>2 mg</td> </tr> <tr> <td>0.02 g</td> <td>0.02 g</td> </tr> <tr> <td>2 g</td> <td>2 g</td> </tr> <tr> <td>0.02 kg</td> <td>0.02 kg</td> </tr> <tr> <td>0.2 kg</td> <td>0.2 kg</td> </tr> </tbody> </table>	<u>Lab</u>	<u>On-site</u>	0.2 mg	0.3 mg	2 mg	2 mg	0.02 g	0.02 g	2 g	2 g	0.02 kg	0.02 kg	0.2 kg	0.2 kg
<u>Lab</u>	<u>On-site</u>																
0.2 mg	0.3 mg																
2 mg	2 mg																
0.02 g	0.02 g																
2 g	2 g																
0.02 kg	0.02 kg																
0.2 kg	0.2 kg																

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 12 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
<p>4. Torque Wrench & Hand Torque Tools</p> <p>Type I & Type II</p> <p>1 to < 2 Nm 2 to < 10 Nm 10 to < 50 Nm 50 to < 100 Nm 100 to < 1000 Nm 1000 to < 2500 Nm</p>	<p>In-house procedure HYL/WI/M-10 Rev: 03 & ISO 6789:2017</p>	<p><u>(Lab only)</u></p> <p>3.2 % Rdg 1.6 % Rdg 2.1 % Rdg 1.6 % Rdg 1.1 % Rdg 1.8 % Rdg</p>
<p>5. Pressure Vacuum</p> <p>a) Vacuum Instruments Analog gauge, Digital gauge/ Calibrators, Manometer, Switch</p> <p><u>Range :</u> -0.95 bar to 0 bar</p>	<p>In-house procedure HYL / WI / M-02, Rev: 04, HYL / WI / M-04, Rev: 04,</p>	<p><u>(Lab / On-Site)</u> 0.0043 bar</p>
<p>b) Pneumatic Pressure Instruments</p> <p>Analog gauge, Digital gauge/ Calibrators, Manometer, Switch</p> <p><u>Range :</u> 0 to 700 mbar 0.7 bar to 7 bar 7 bar to 35 bar</p>	<p>In-house procedure HYL / WI / M-02, Rev: 04, HYL / WI / M-04, Rev: 04,</p>	<p><u>(Lab / On-Site)</u> 0.8 mbar 0.02 bar 0.052 bar</p>

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 13 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
<p>c) Hydraulic Pressure Instruments Analog Gauge, Digital Gauge/ Calibrators, Manometer, Switch</p> <p><u>Range :</u> 36 bar to 70 bar 71 bar to 350 bar 351 bar to 700 bar 701 bar to 1000 bar 1000 bar to 1400 bar</p>	<p>In-house procedure HYL / WI / M-02, Rev: 04, HYL / WI / M-04, Rev: 04, By DPC</p>	<p>0.11 bar 0.2 bar 0.4 bar 0.9 bar 1.7 bar</p>
<p>d) Pressure Transmitter</p> <p><u>Range :</u> -0.95 bar to 3 bar Above 3 bar to 100 bar Above 100 bar to 600 bar</p>	<p>In-house procedure HYL / WI / M-03 Rev: 02</p>	<p><u>(Lab / On-Site)</u> 0.01 bar 0.16 bar 1.0 bar</p>
<p>e) Hydraulic Pressure Instruments Analog Gauge, Digital Gauge /Calibrators, Manometer, Switch</p> <p><u>Range :</u> 1 bar to 30 bar 30 bar to 70 bar 71 bar to 700 bar (The range 701 bar to 1300 bar scope was removed due to lab facility limit)</p>	<p>In-house procedure HYL / WI / M-02, Rev: 04, HYL / WI / M-04, Rev: 04, By DWT</p> <p>HYL / WI / M-02, Rev: 04, HYL / WI / M-04, Rev: 04, By DWT</p>	<p><u>(Lab only)</u> 0.01 bar 0.02 bar 0.03 % Rdg</p>

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 14 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)																		
<p>C. TEMPERATURE</p> <p>1. Calibration of RTD</p> <p><u>Range:</u> -75 °C to -10 °C >-10 °C to 50 °C > 50 °C to 250 °C > 250 °C to 350 °C > 350 °C to 650 °C</p> <p>2. Calibration of Liquid Bath</p> <p><u>Range:</u> -75 °C to 50 °C > 50 °C to 250 °C</p> <p>3. Calibration of Dry Block</p> <p><u>Range:</u> -50 °C to 0 °C > 0 °C to 50 °C > 50 °C to 200 °C > 200 °C to 350 °C > 350 °C to 650 °C</p> <p>4. Calibration of Digital Thermometer for RTD -Simulate & Measure</p> <p><u>Range:</u> -200 °C to 850 °C (Measure) -200 °C to 850 °C (Simulate)</p>	<p>In-house procedure HYL / WI / T-02 Rev: 06</p> <p>In-house procedure HYL / WI / T-06 Rev: 03</p> <p>In-house procedure HYL / WI / T-04 Rev: 05</p> <p>In-house procedure HYL / WI / T-07 Rev: 03</p>	<table border="0"> <tr> <td style="text-align: center;"><u>Lab</u></td> <td style="border-left: 1px solid black; text-align: center;"><u>On-site</u></td> </tr> <tr> <td style="text-align: center;">0.3 °C</td> <td style="border-left: 1px solid black; text-align: center;">N.A.</td> </tr> <tr> <td style="text-align: center;">0.3 °C</td> <td style="border-left: 1px solid black; text-align: center;">N.A.</td> </tr> <tr> <td style="text-align: center;">0.5 °C</td> <td style="border-left: 1px solid black; text-align: center;">0.5 °C</td> </tr> <tr> <td style="text-align: center;">0.6 °C</td> <td style="border-left: 1px solid black; text-align: center;">0.6 °C</td> </tr> <tr> <td style="text-align: center;">0.8 °C</td> <td style="border-left: 1px solid black; text-align: center;">0.8 °C</td> </tr> </table> <p style="text-align: center;"><u>(Lab / On-Site)</u></p> <p style="text-align: center;">0.3 °C 0.4 °C</p> <p style="text-align: center;"><u>(Lab / On-Site)</u></p> <p style="text-align: center;">0.2 °C 0.2 °C 0.5 °C 0.6 °C 0.7 °C</p> <table border="0"> <tr> <td style="text-align: center;"><u>(Lab)</u></td> <td style="border-left: 1px solid black; text-align: center;"><u>(On-Site)</u></td> </tr> <tr> <td style="text-align: center;">0.11 °C</td> <td style="border-left: 1px solid black; text-align: center;">0.30 °C</td> </tr> <tr> <td style="text-align: center;">0.04 °C</td> <td style="border-left: 1px solid black; text-align: center;">0.34 °C</td> </tr> </table>	<u>Lab</u>	<u>On-site</u>	0.3 °C	N.A.	0.3 °C	N.A.	0.5 °C	0.5 °C	0.6 °C	0.6 °C	0.8 °C	0.8 °C	<u>(Lab)</u>	<u>(On-Site)</u>	0.11 °C	0.30 °C	0.04 °C	0.34 °C
<u>Lab</u>	<u>On-site</u>																			
0.3 °C	N.A.																			
0.3 °C	N.A.																			
0.5 °C	0.5 °C																			
0.6 °C	0.6 °C																			
0.8 °C	0.8 °C																			
<u>(Lab)</u>	<u>(On-Site)</u>																			
0.11 °C	0.30 °C																			
0.04 °C	0.34 °C																			

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 15 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)	
<p>5. Calibration of Digital Thermometer for Thermocouple</p> <p><u>Range:</u> Type J (Measure): -200 °C to 1200 °C Type J (Simulate): -200 °C to 1200 °C Type K (Measure): -260 °C to 1370 °C Type K (Simulate): -260 °C to 1370 °C Type T (Measure): -260 °C to 400 °C Type T (Simulate): -260 °C to 400 °C Type N (Measure): 0 °C to 1200 °C Type N (Simulate): 0 °C to 1200 °C Type E (Measure): -200°C to 1000°C Type E (Simulate): -200°C to 1000°C Type R (Measure): 170 °C to 1700°C Type R (Simulate): 170 °C to 1700°C Type S (Measure): 170 °C to 1700°C Type S (Simulate): 170 °C to 1700°C</p>	<p>In-house procedure HYL / WI / T-08 Rev: 05</p>	<p>(Lab)</p> <p>0.52 °C 0.40 °C 0.53 °C 0.41 °C 0.58 °C 0.39 °C 0.51 °C 0.46 °C 0.45 °C 0.42 °C 1.6 °C 1.4 °C 1.5 °C 1.4 °C</p>	<p>(On-Site)</p> <p>0.53 °C 0.40 °C 0.53 °C 0.42 °C 0.60 °C 0.40 °C 0.52 °C 0.47 °C 0.46 °C 0.43 °C 1.6 °C 1.4 °C 1.5 °C 1.4 °C</p>
<p>6. Calibration of Temperature Transmitter</p> <p><u>Range:</u> - 50 °C to 50 °C > 50 °C to 300 °C > 300 °C to 600 °C</p>	<p>In-house procedure HYL / WI / T-09 Rev: 02</p>	<p>(Lab / On-Site)</p> <p>0.2 °C 0.2 °C 0.3 °C</p>	
<p>7. Calibration of Mechanical Temperature Gauges / Bimetal Thermometer / Surface Probe</p> <p><u>Range:</u> - 50 °C to 0 °C > 0 to 50 °C > 50 °C to 200 °C > 200 °C to 600 °C</p>	<p>In-house procedure HYL / WI / T-02 Rev: 06</p>	<p>(Lab)</p> <p>0.4 °C 0.5 °C 0.5 °C 1.0 °C</p>	<p>(On-Site)</p> <p>NA NA 0.5 °C 1.0 °C</p>

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 16 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
<p>8. Calibration of Temperature and Humidity measuring instruments</p> <p><u>Range:</u> 15 °C to 45 °C @ 55 %Rh* 30 %Rh to 80 % Rh @ 25 °C * %Rh= % Relative humidity</p>	<p>In-house procedure HYL / WI / T-11 Rev: 02</p>	<p><u>(Lab)</u> 0.5 °C 2.7 %Rh</p>
<p>9. Calibration of Humidity Chamber</p> <p><u>Range:</u> 15 °C to 45 °C 30 %Rh to 80% Rh</p>	<p>In-house procedure HYL / WI / T-10 Rev: 02</p>	<p><u>(Lab / On-Site)</u> 1.4 °C 9.4%Rh</p>
<p>10. Calibration of Temperature Enclosure / Oven / Freezer / Incubator</p> <p><u>Range:</u> -50 °C to 50 °C 50 °C to 200 °C 200 °C to 700 °C</p> <p>Up to 125 °C (Pressurized vessel / Autoclave)</p>	<p>In-house procedure HYL / WI / T-10 Rev: 03</p>	<p><u>(Lab / On-Site)</u> 1.8 °C 2.8 °C 5.8 °C 0.61 °C</p>

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 17 of 18

MEASURED QUANTITIES/ INSTRUMENT/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC *)
D. 1. Dimensional (Lab Only) External Micrometer (Fixed Anvil) Range : 0 – 25mm Resolution: 0.01mm 0.001 mm 0.01mm 0.001mm Note: (#) Range of Error of Traverse of Micrometer Screw	In-house procedure HYL / WI / M-18 Rev: 01, 8 June 2013 Set of Slip Gauges	6 µm 1 µm 9 µm # 2 µm #
2. Vernier Caliper (Outside/Inside) Range : 20 mm – 50 mm 50 mm – 350 mm 350 mm – 400 mm	In-house procedure HYL / WI / M-17 Rev: 01, 8 June 2013 By Caliper Checker	13 µm 14 µm 15 µm
3. Digital Caliper (Outside/Inside) Range : 20 mm - 250 mm 250 mm - 400 mm	In-house procedure HYL / WI / M-17 Rev: 01, 8 June 2013 By Caliper Checker	9 µm 10 µm
4. Dial Caliper (Outside/Inside) <u>Range :</u> 20 mm - 250 mm 250 mm - 300 mm	In-house procedure HYL / WI / M-17 Rev: 01, 8 June 2013 By Caliper Checker	9 µm 10 µm

* CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95%.

Schedule



Certificate No. : LA-2013-0549-C

Issue No. : 9

Date : 12 December 2022

Page : 18 of 18

Approved signatories :

Mr Steve Eer) For all categories

Mr Saranraj Tamilarasan) For all items in Section A, B and C

Mr. Muhammad Haqem Rafiq Geach) For all items in Section B

Note :

This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibration results. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.